

CE

**TECHNOLOGIES** 



LT/RC-179000-A+B/M

### ASTM D323 ASTM D4953 IP 69 ISO 3007

## ASTM D323 - IP 69 - ISO 3007

### Vapour Pressure of Petroleum Products (Reid Method) Procedure B

The test method covers the determination of vapour pressure of gasoline, volatile crude oil, by means of three procedures: A, B and C.

The OilLab 715 grant the determination in conformity with the procedure B (horizontal bath) on petroleum products having Reid Vapour Pressures below 180 kPa (26 psi).

### ASTM D4953

# Standard Test Method for Vapor Pressure of Gasoline and Gasoline-Oxygenate Blends

This test method covers and is applicable to gasolines and gasoline-oxygenate blends with a vapor pressure range from 35 kPa

to 100 kPa (5 psi to 15 psi).

This test method, a modification of Test Method D323 (Reid Method), provides

two procedures to determine the vapor pressure of gasoline and gasoline-oxygenate blends.

- Bath made in stainless steel suitable for the accommodation of up to two (2) vapour pressure cylinders according to ASTM specifications.
- Bath temperature range from ambient to +80°C  $\pm 0.1^{\circ}$ , filling level 20 litres.
- Analysis range from 0 up to 250 kPa.
- Motorized stirrer with shaft, drain tap
  Electric immersion Heater controlled by PID system.
- Drive system for swing gently the vessel in horizontal position.
- Secure handle cover for open bath.
- Bath equipped with a Touch screen panel PC: TFT/LCD 8";
  - · 40 Gb HD;
- · Resolution 1024 × 768 and 16M colours;
- $\cdot$  2  $\times$  USB Ports for connecting pen drive and printer.
- The dedicated software manages:
- The bath temperatures by means of a PT100 sensor class;
  A that can be displayed in °C / °F,
- including the over temperature safety alarm.
  2 (two) electronic sensors, pressure transducers / electronic pressure gauges supplied for the
- coupling to the test vessels. • Cables and connectors.
- Dedicated software for real time monitoring and recording that includes:
- Display of the pressure in bar / psi / Kpa;
  Graph creation in real time during the test;
- Invalid test indication;
- Export of files in xls / pdf / jpg format;
- Calibration up to 100 points.
- Power supply: 220Vac 50/60Hz.

#### Accessories

- LAB-100-371/50: silicone oil, can of 25 litres
- T-AS18C: thermometer ASTM 18C
- LAB-102-013: flexible Junction for O<sub>2</sub>
- LT/RC-179000/M:
- Reid Cylinder Liquid Chamber One Opening LT/RC-179000-A/M - ASTM D323
- · made in stainless steel
- · in one end of the liquid chamber
- an opening of approximately ½" shall be provided for coupling with the vapour
- chamber • the inner surface of the coupling end shall be sloped to provide complete drainage when
- inverted the other end of the chamber shall
- be completely closed • Reid Cylinder - Vapour Chamber
- LT/RC-179000-B/M ASTM D323
- $\cdot\,\text{made}$  in stainless steel
- · lower coupling
- $\cdot$  upper ½" groove with a ¼" reducing cap for pressure gauge